



		27-33 kV	24-48 kV	01-48 kV
4-200mA AND PHOTO TIMER	Focus 0.1	25 mA	25 mA	25 mA
WITH SCREEN	Focus 0.3	100 mA	100 mA	70 mA
250-500mA AND PHOTO TIMER	Focus 0.3	50 mA	50 mA	50 mA
WITHOUT SCREEN				

**WARNING**  
THIS X RAY UNIT MAY  
AND OPERATOR INJURY  
AND OPERATING INSTRU

Mo Al



Two columns of controls. The left column has a small square indicator above a slider switch, and below it a switch with a vertical bar symbol. The right column has a small square indicator above a slider switch, and below it a switch with a circle and tilde symbol.

kV



mA

mAs

# SENOGRAPHE 600T



<b>MODEL NUMBER</b>	Z.1B2B1.4A51	<b>TYPE</b>	39558014
<b>MANUFACTURED</b>	JUNE 1995		
<b>LOCATION</b>	M		
<b>HOUSING</b>	STATORIX M51		
<b>S. N.</b>	11367CX0		
<b>NOM. VOLTAGE</b>	49 kV		
<b>INSERT</b>	GS 512-4		
<b>S. N.</b>	9220TX5		
<b>NOM. VOLTAGE</b>	49 kV		
<b>HOUSING ASSY.</b>			
<b>NOM. VOLTAGE</b>	49 kV		
<b>INH. FILTRATION</b>	0,8 mm Be		
<b>FOCAL SPOTS</b>	0,1-0,3 mm		

